



Prestan professional training manikin

Rate monitor

The Prestan Professional Training Manikin Rate Monitor has a unique way of monitoring both the depth of compressions and the rate of compressions.

At the first compression all four of the lights in the upper left shoulder will come on, indicating that the lights are working and the batteries are good. (Figure #1)

If the lights do not light, make sure that the depth of compression is deep enough (you will hear a clicking sound). The lights will not work if the student is not compressing the manikin correctly.

If the depth of compression is correct, and the lights do not come on, replace the two AA alkaline batteries in the battery compartment located in the back of the manikin in the left shoulder.

Once a student begins chest compressions, the amber light and the green lights will go out and, from zero to less than 60 compressions per minute, the red light will be on. (Figure #2)

As the student increases the rate of compressions and reaches 60 compressions per minute, the red light will go off and the amber light will come on until the student reaches 80 compressions per minute. (Figure #3)

As the student increases the rate of compressions and reaches 80 compressions per minute, the amber light will go off and the first green light will come on until the student reaches 100 compressions per minute. (Figure #4)

When the student reaches 100 compressions per minute, the second green light comes on, indicating that the proper rate has been attained. (Figure #5)

If the student slows down, one of the green lights will go out indicating that the student needs to increase the rate of compressions.

If the student's depth of compression is not correct, at any time, all the lights will go off and the red light will blink.



FIGURE #1



FIGURE #1



FIGURE #1



FIGURE #1



FIGURE #5